

FRIEDMAN & BRUYA, INC.

ENVIRONMENTAL CHEMISTS

Date of Report: 05/26/04

Date Received: 05/19/04

Project: Metro KC Comp., PO# M104246, F&BI 405183

Date Extracted: 05/21/04

Date Analyzed: 05/21/04

**RESULTS FROM THE ANALYSIS OF THE WATER SAMPLE
FOR TOTAL METALS BY
INDUCTIVELY COUPLED PLASMA (ICP)
(METHOD 6010)**

Results Reported as mg/L (ppm)

<u>Sample ID</u> Laboratory ID	<u>Chromium</u>	<u>Copper</u>	<u>Nickel</u>	<u>Zinc</u>
M104246 405183-01	0.50	0.55	0.60	<0.05
Method Blank	<0.05	<0.05	<0.10	<0.05

FRIEDMAN & BRUYA, INC.**ENVIRONMENTAL CHEMISTS**

Date of Report: 05/26/04

Date Received: 05/19/04

Project: Metro KC Comp., PO# M104246, F&BI 405183

**QUALITY ASSURANCE RESULTS
FROM TOTAL METALS BY
INDUCTIVELY COUPLED PLASMA (ICP)
(METHOD 6010)**

Laboratory Code: 405183-01 (Duplicate)

Analyte	Reporting Units	Sample Result	Duplicate Result	Relative Percent Difference	Acceptance Criteria
Chromium	mg/L (ppm)	0.50	0.50	0	0-20
Copper	mg/L (ppm)	0.55	0.56	2	0-20
Nickel	mg/L (ppm)	0.60	0.63	5	0-20
Zinc	mg/L (ppm)	<0.05	<0.05	nm	0-20

Laboratory Code: 405183-01 (Matrix Spike)

Analyte	Reporting Units	Spike Level	Sample Result	% Recovery MS	Acceptance Criteria
Chromium	Mg/L (ppm)	2	0.50	89	80-120
Copper	mg/L (ppm)	2	0.55	90	80-120
Nickel	mg/L (ppm)	4	0.60	91	80-120
Zinc	mg/L (ppm)	2	<0.05	95	80-120

Laboratory Code: Laboratory Control Sample

Analyte	Reporting Units	Spike Level	% Recovery LCS	Acceptance Criteria
Chromium	mg/L (ppm)	2	95	80-120
Copper	mg/L (ppm)	2	93	80-120
Nickel	mg/L (ppm)	4	95	80-120
Zinc	mg/L (ppm)	2	96	80-120

nm - The analyte was not detected in one or more of the duplicate analyses. Therefore, calculation of the RPD is not applicable.

405183

SAMPLE CHAIN OF CUSTODY

CM 05-19-04

BI,

Send Report To

~~ALASKAN~~ G. Thompson

Company

ALASKAN Copper Works

Address

628 S. Hancock St

City, State, ZIP

Seattle WA 98134

Phone #

206-882-8379

Fax #

206-882-4309

SAMPLERS (signature)

PROJECT NAME/NO.

METRO KC Comp.

PO #

M 104246

REMARKS

Page # of

TURNAROUND TIME

☐ Standard (2 Weeks)☒ RUSH 24 hrs

Rush charges authorized by:

SAMPLE DISPOSAL

☐ Dispose after 30 days☐ Return samples☐ Will call with instructions

Sample ID	Lab ID	Date	Time	Sample Type	# of containers	ANALYSES REQUESTED										Notes
						TPH-Diesel	TPH-Gasoline	BTEX by 8021B	VOCs by 8260	SVOCs by 8270	HFS					
M/104246	01	5/19/04	11:00	A/C	1											

Friedman & Bruya, Inc.
3012 16th Avenue West

Seattle, WA 98119-2029

Ph. (206) 285-8282

Fax (206) 283-5044

FORMS\COC\COC.DOC

SIGNATURE

Relinquished by:

Received by:

Relinquished by:

Received by:

PRINT NAME

G. Thompson
Nhan Phan

COMPANY

ACW
FBI

DATE

5/15/04
5/17/04

TIME

1:23pm
1:23pm

FRIEDMAN & BRUYA, INC.

ENVIRONMENTAL CHEMISTS

James E. Bruya, Ph.D.
Charlene Morrow, M.S.
Yelena Aravkina, M.S.
Bradley T. Benson, B.S.
Kurt Johnson, B.S.

3012 16th Avenue West
Seattle, WA 98119-2029
TEL: (206) 285-8282
FAX: (206) 283-5044
e-mail: fbi@isomedia.com

DUPLICATE COPY

May 26, 2004

INVOICE #04ACU0526-1

Accounts Payable
Alaskan Copper Works
628 South Hanford
Seattle, WA 98134

RE: Project Metro KC Comp., PO# M104246, F&BI 405183 - Results of testing
requested by Gerry Thompson for material submitted on May 19, 2004.

1 sample analyzed for Total Chromium, Copper, Nickel and Zinc by Method 6010B @ \$80 per sample	\$ 80.00
Amount Due	\$ 80.00

FEDERAL TAX ID #(b) (6)

FRIEDMAN & BRUYA, INC.

ENVIRONMENTAL CHEMISTS

James E. Bruya, Ph.D.
Charlene Morrow, M.S.
Yelena Aravkina, M.S.
Bradley T. Benson, B.S.
Kurt Johnson, B.S.

3012 16th Avenue West
Seattle, WA 98119-2029
TEL: (206) 285-8282
FAX: (206) 283-5044
e-mail: fbi@isomedia.com

May 26, 2004

Gerry Thompson, Project Manager
Alaskan Copper Works
628 South Hanford
Seattle, WA 98134


Dear Mr. Thompson:

Included are the results from the testing of material submitted on May 19, 2004 from the Metro KC Comp., PO# M104246, F&BI 405183 project. There are 2 pages included in this report. Any samples that may remain are currently scheduled for disposal in 30 days. If you would like us to return your samples or arrange for long term storage at our offices, please contact us as soon as possible.

We appreciate this opportunity to be of service to you and hope you will call if you should have any questions.

Sincerely,

FRIEDMAN & BRUYA, INC.



Charlene Morrow
Chemist

Enclosures
ACU0526R.DOC